# Certificate of Accreditation



# Spline Gauges A trading name of Apex Tool Group (UK Operations) Ltd

Calibration Laboratory No. 0015

Is accredited in accordance with International Standard ISO/IEC 17025:2017 – General Requirements for the competence of testing and calibration laboratories.

This accreditation demonstrates technical competence for a defined scope specified in the schedule to this certificate, and the operation of a management system (refer joint ISO-ILAC-IAF Communiqué dated April 2017). The schedule to this certificate is an essential accreditation document and from time to time may be revised and reissued.

The most recent issue of the schedule of accreditation, which bears the same accreditation number as this certificate, is available from www.ukas.com.

This accreditation is subject to continuing conformity with United Kingdom Accreditation Service requirements.

Matt Gantley, Chief Executive Officer United Kingdom Accreditation Service

Initial Accreditation: 6 March 1969 Certificate Issued: 3 March 2020







Scan QR Code to verify

## **Schedule of Accreditation**

issued by

# **United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0015

Accredited to ISO/IEC 17025:2017

**B78 2ER** 

# Spline Gauges A trading name of Apex Tool Group (UK Operations) Ltd

Issue No: 021 Issue date: 15 September 2021

 Piccadilly
 Contact: Mr Gareth Jones

 Tamworth
 Tel: +44 (0)1827-872771

 Staffordshire
 Fax: +44 (0)1827-874128

E-Mail: gareth.jones@apextoolgroup.com

Website: www.splinegauges.co.uk

#### Calibration performed at the above address only

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range		Expanded Measurement Uncertainty (k = 2)	Remarks
RANGE IN MILLIMETRES	S AND UNCERTA	AINTY IN MIC	ROMETRES UNLESS OTHERV	VISE STATED
INVOLUTE GEARS, GEAR ARTEFACTS, SPLINE GAUGES (see notes 1 and 2)				NOTES
External				Gears of the following capacities may be calibrated:     Maximum diameter 150 mm,     Maximum length 100 mm,     Max Weight 30 kg
				2. The uncertainties stated assume that journal diameters or reference surfaces have been used to define the measurement axis.
Profile Total deviation $(F_{\alpha})$ Profile slope deviation $(f_{H\alpha})$ Profile form deviation $(f_{f\alpha})$	-		1.4 1.3 1.4	CNC gear measuring machine.
Helix (Alignment) Total deviation $(F_\beta)$ Helix (alignment) slope deviation $(f_{H\beta})$ Helix (alignment) form deviation $(f_{\beta f})$	Helix angle		1.6 1.5 1.6	
Single Pitch (fp)	0° to 45°		1.8	
Pitch Difference (f <sub>u</sub> )		0.15 to 25 Module	1.8	
Cumulative Pitch (F <sub>p</sub> )			2.7	
Radial Runout of Tooth Space (F <sub>r</sub> )			3.3	
Normal Circular Tooth Thickness(S <sub>n</sub> )			1.6	Horizontal measuring machine
Dimension Over/Pins or Balls (Mdr or Mdk)	5 to 100 100 to 200 200 to 250 250 to 300	<b>⊥</b>	1.7 1.9 2.1 2.9	and reference setting standards.

Assessment Manager: CD1 Page 1 of 5



# Schedule of Accreditation issued by United Kingdom Accréditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Spline Gauges A trading name of Apex Tool Group (UK Operations) Ltd

Issue date: 15 September 2021 Issue No: 021

#### Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range		Expanded Measurement Uncertainty (k = 2)	Remarks		
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED						
INVOLUTE GEARS, GEAR ARTEFACTS, SPLINE GAUGES (see notes 1 and 2) (cont'd)				NOTES (cont'd)		
Internal						
Profile Total deviation $(F_{\alpha})$ Profile slope deviation $(f_{H\alpha})$ Profile form deviation $(f_{f\alpha})$			1.4 1.3 1.4	CNC gear measuring machine.		
Helix (Alignment) Total deviation ( $F_{\beta}$ ) Helix (alignment) slope deviation ( $f_{H\beta}$ ) Helix (alignment) form deviation ( $f_{\beta f}$ )	Helix angle  0° to 45°		1.6 1.5 1.6			
Single Pitch (f <sub>p</sub> )		0.15 to 25	1.8			
Pitch Difference (f <sub>u</sub> )		Module	1.8			
Cumulative Pitch (F <sub>p</sub> )			2.7			
Radial Runout of Tooth Space (F <sub>r</sub> )			3.3			
Normal Circular Tooth Thickness( $S_n$ )			1.6			
Dimension Between Pins or Balls	5 to 100 diameter		2.6	Horizontal measuring machine and reference setting standards.		
(Mdr or Mdk)	100 to 200 diameter		2.9			
STRAIGHT SIDED SERRATION GAUGES						
Plug						
Serration Angle Dimension Across Flats 90° Only.	_		10 Minutes of Arc 1.6	CNC gear measuring machine or Con-tracer.		
Single Pitch (f <sub>p</sub> )		0.15 to 25	1.8	CNC gear measuring machine.		
Pitch Difference (f <sub>u</sub> )		Module	1.8			
Cumulative Pitch (F <sub>p</sub> )			2.7			
Dimension Over Pins or Balls (Mdr or Mdk)			1.7	Horizontal measuring machine and reference setting standards.		
Straight Sided Plug Tooth Thickness			1.2			

Assessment Manager: CD1 Page 2 of 5



# Schedule of Accreditation issued by United Kingdom Accréditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Spline Gauges A trading name of Apex Tool Group (UK Operations) Ltd

Issue No: 021 Issue date: 15 September 2021

#### Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range		Expanded Measurement Uncertainty (k = 2)	Remarks
RANGE IN MILLIMETRE	S AND UNCER	TAINTY IN MIC	ROMETRES UNLESS OTHER	RWISE STATED
STRAIGHT SIDED SERRATION GAUGES (cont'd)				
Ring				
Serration Angle Dimension Across Flats 90° Only.	7		10 Minutes of Arc 1.6	CNC gear measuring machine or Contracer.
Single Pitch (f <sub>p</sub> )			1.8	CNC gear measuring machine.
Pitch Difference (f <sub>u</sub> )		0.15 to 25 Module	1.8	
Cumulative Pitch (F <sub>p</sub> )		iviouuie	2.7	
Dimension Between Pins or Balls (Mdr or Mdk)	20 to 100 D	Piameter	2.6	Horizontal measuring machine and reference setting
Straight Sided Internal Tooth Thickness			1.2	standards.
GENERAL				
Bore Diameters	5 to 25 Diam	eter	1.0	Horizontal measuring machine and reference setting
	25 to 50 Diar	meter	1.3	standards.
	50 to 100 Dia	ameter	1.9	
Major Diameter (Even Teeth)	5 to 100 Diar	meter	1.1	
	100 to 175 D	iameter	1.5	
	175 to 250 Diameter		1.9	
	250 to 300 D	iameter	2.3	
Major Diameter (Odd Teeth)	5 to 100 Diameter		1.7	
	100 to 175 Diameter		1.9	
	175 to 250 Diameter		2.2	
	250 to 300 D	iameter	2.4	
Minor Diameter (Even Teeth)	Teeth) 10 to 100 Diameter		1.1	
	100 to 200 D	iameter	1.9	
Minor Diameter (Odd Teeth)	10 to 100 Diameter 100 to 200 Diameter		1.7	
			2.2	

Assessment Manager: CD1 Page 3 of 5



# Schedule of Accreditation issued by United Kingdom Accréditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Spline Gauges A trading name of Apex Tool Group (UK Operations) Ltd

Issue No: 021 Issue date: 15 September 2021

#### Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
GENERAL (cont'd)			NOTES (cont'd)	
Chamfer		12.7	Con-tracer.	
Fillet radius		12.7	Con-tracer.	
Radial and axial runout		2.5	CNC gear measuring machine	
END				

Assessment Manager: CD1 Page 4 of 5



# Schedule of Accreditation issued by United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

# Spline Gauges A trading name of Apex Tool Group (UK Operations) Ltd

Issue No: 021 Issue date: 15 September 2021

#### Calibration performed at main address only

#### Appendix - Calibration and Measurement Capabilities

#### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

#### **Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k = 2. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

#### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where q is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets: Q[a, b] =  $[a^2 + b^2]^{1/2}$ 

Assessment Manager: CD1 Page 5 of 5